


ASD Documentation Received - 100 Mile House

Thursday, February 28, 2013
11:42 AM

Subject	ASD Documentation Received - 100 Mile House
From	McKeown, Kayla JAG:EX
To	Burchnall, Shelly K JAG:EX; Anderson, Kathy E JAG:EX; Esposito, Tony N JAG:EX
Sent	Thursday, July 5, 2012 4:51 PM
Attachments	 2012070516 4547

Hi Everyone,

Attached is a document received from 100 Mile House RCMP in relation to a police supplemental report regarding an ASD. I have checked through all of the processed IRP files from 100 Mile House and below is the information for the one IRP served that used the particular ASD in question. I will check through the files again on Monday and Tuesday next week to ensure that all IRPs processed have covered all appropriate dates for the ASD.

Please let me know if you require any other information.


Thanks!

Date of Offence	Driver Name	DL#	Prov/St ate	Licence Class	Prohibiti on Type	IRP/ADP Notice #	First ASD Result	First ASD Serial #	Second ASD Result	Second ASD Serial #	VI Notice #	VI Offence	VI Length	Police File Number	Police Detach ment
s.22			BC	560	90 day IRP - FAIL	s.22	Not Responsive		FAIL	100835	s.22	30 day impoun dment (30 or 90 day IRP)	30 day	s.22	100 Mile House Traffic Service s (Caribo o - Chilcoti n Region al Traffic Unit)

Kayla McKeown
Team Lead, Branch Support & Manual Operations
Administrative Justice Branch
The Office of the Superintendent of Motor Vehicles | Ministry of Justice
PO Box 9254 STN Prov Govt | Victoria BC | V8W 9J2 | 250-356-5634



From: Kayla.McKeown@gov.bc.ca [mailto:Kayla.McKeown@gov.bc.ca]
Sent: Thursday, July 5, 2012 4:46 PM
To: McKeown, Kayla JAG:EX
Subject:



POLICE SUPPLEMENTAL TO FILE REPORT

Document Type: ☒ IRP _____ IRP No. _____
☐ VI _____ VI No. _____
☐ UL _____ UL No. _____
☐ ADP _____ ADP No. _____
☐ 215 (24hr) _____ 215 No. _____

Driver's Surname		Driver's Given Name	Initials
D.O.B.	B.C. Drivers License No.	Police File No.	
Detachment 100 Mile House - RCMP			
Issuing Officer's Surname HUTT		Issuing Officer's Given Name Alex . D	Telephone No. 250 395-2456

Please be advised that an Alco-Sensor IV DWF Accuracy Check dated July 3rd, 2012 determined that Alco-Sensor IV DWF serial number 100835 was providing falsely high readings. The date of the preceding acceptable Accuracy Check was June 19, 2012. This device has been sent to the service agent for maintenance or repair.

2012-07-03
Date

FILING INSTRUCTIONS Completed original to operational file.
Copy to the Office of the Superintendent of Motor Vehicles.




Canada

ASD ACCURACY CHECK LOG

Simulator SIN:	6 9289
True-Cal SIN:	00 7777

ASD SIN: AS IV 100835
Location: 100 Pine House

Alc. Std.	(Dry Gas) Value @ Sea level <i>0.21 mg %</i>	(Skin Sol'n) Expected value: <i>100 mg %</i>
Info.	Manufacturer: <i>AIRGAS</i>	Manufacturer: <i>CORNING</i>

Date	CALIBRATION CHECK				RECALIBRATION		CALIBRATION CHECK				BLANK TEST		ASD Calibrator Signature	
	Dry Gas Air Std Lot #	Lot Exp Date	True-Cal Expected Value	Calibration Check mg/100mL	Wet Bath Air Std Lot #	Lot Exp Date	# Samples taken	True-Cal Expected Value	Calibration Check mg/100mL	True-Cal Expected Value	Calibration Check mg/100mL	Blank Test mg/100mL		
15 Jun 2012	06-118802	07 Jul 2013	0.73	0.73	20723	2013/12	14	0.73	0.77	0.73	0.75	0	19 Jun 2012	
19 Jun 2012	06-118802	07 Jul 2013	0.73	0.70								0	03 Jul 2012	
23 Jul 2012	06-118802	07 Jul 2013	0.73	0.116				REQUIRED TO DATECH						

06/25/12 09:24:57 AL VICTORIA OSMVOSMV

linked, stats - alistair

Pages 5 through 10 redacted for the following reasons:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2013-05-15

ASD SERIAL NUMBER
100835

CERTIFICATE OF QUALIFIED ASD CALIBRATOR

I, Harold G. DEAN, certify that:

I am a qualified ASD Calibrator, qualified to calibrate the ☒ Alco-Sensor IV DWF
and (check one): ☐ Alco-Sensor FST

☒ On the 19th day of June, 2012, I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: AIRGAS Lot Number: AG 118802, Expiry: 7-JUL-2013.

It was found to be within the recommended limits.

OR

☐ On the _____ day of _____, 20____, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____ and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of July 3rd, 2012.

Dated this 19th day of June, 2012, at 100 Mile House in the Province of British Columbia

Calibrator: Harold G. DEAN


Dept/Reg Number: E-2196 / 110099

Agency: 100 Mile House - RCMP


Calibrator Signature

ASD Documentation Received - Clinton RCMP

Thursday, February 28, 2013
11:41 AM

Subject	ASD Documentation Received - Clinton RCMP
From	McKeown, Kayla JAG:EX
To	Anderson, Kathy E JAG:EX; Esposito, Tony N JAG:EX
Sent	Thursday, September 20, 2012 2:53 PM
Attachments	 2012092013 3259

Hi Everyone,

Attached is a document received from Clinton RCMP in relation to a police supplemental report regarding an ASD. I have checked through all of the processed IRP files from Clinton and below is the information for the one IRP served that used the particular ASD in question. I will check through the files again on Monday and Tuesday next week to ensure that all IRPs processed have covered all appropriate dates for the ASD.

Please let me know if you require any other information.

Thanks!

Driver's Name	ADP/IR P Prohibition Notice Number	ADP/IR P Offence Date	ADP/IR P Prohibition Type	ADP/IR P MVA Section	ADP/IR P Police File Number	ADP/IR P Police Detachment	ADP/IR P First ASD Result	ADP/IR P First ASD Serial Number	ADP/IR P Second ASD Result	ADP/IR P Second ASD Serial Number	VI Notice Number	VI Police File Number	VI Police Detachment
s.22			IRP 90 day	ASD Fail	s.22	Clinton	Not Responsive		FAIL	67783	s.22		Clinton

Kayla McKeown
Team Lead, Branch Support & Manual Operations
Administrative Justice Branch
The Office of the Superintendent of Motor Vehicles | Ministry of Justice
PO Box 9254 STN Prov Govt | Victoria BC | V8W 9J2 | 250-356-5634



Page 13 redacted for the following reason:

S.15, s.16, s.22

September 27, 2012

s.22

Immediate Roadside Prohibition (IRP) No. s.22

I am writing with respect to the Notice of Driving Prohibition served on you dated August 19, 2012.

As the Notice is invalid, the driving prohibition and its related vehicle impoundment have been cancelled and removed from your driver's licence.

I note that your vehicle has been released. Upon receipt of your proof of payment, the Superintendent will pay towing and storage costs up to and including September 19, 2012, which is the date the vehicle was eligible for release. Please remit a copy of this letter with a copy of the Impound Lot Invoice and your receipt of payment to this office via the address or fax number below. You are responsible for any charges beyond this date.

Appeals Registry
Office of the Superintendent of Motor Vehicles

km

08/22/12 09:12:51 JC VICTORIA OSMVOSMV

Linked stats updated, letter logged, faxed for documents - James

08/24/12 13:52:49 JC VICTORIA OSMVOSMV

Linked - James

09/19/12 15:13:41 AL VICTORIA OSMVOSMV

linked only - alistair

09/26/12 16:50:56 KM VICTORIA OSMVOSMV

police supplemental report received - given to Arlene Sep 20/12 to review.

The report was discussed with Tony and Kayla and the IRP is invalid.

09/27/12 07:54:35 KM VICTORIA OSMVOSMV

A letter has been sent to the client to advise of the invalid IRP due to the ASD inaccuracy. The IRP has been removed from his DL and RDP advised -s.22

The driver is the RO and the vehicle has already been released so the letter also advises him to send in receipts for towing and storage and OSMV will refund to the end of September 19/12 the date of release.

10/16/12 14:40:05 JA VICTORIA OSMVOSMV

Letter dated advising client to send in receipts for towing and storage dated Sept 27/12 in return mail. letter remailed to updated address. Jackie

Pages 16 through 21 redacted for the following reasons:

S.15, s.16, s.22



Ministry of Justice
Office of the Superintendent of Motor Vehicles

URGENT

Enforcement Agency: Clinton

Fax: 250 459 7055 File number: s.22

The following documents (if they exist) are required in our office within 48 hours

**Please include this document with your
return fax to 250-978-8079**

Vehicle Impoundment:

- ☐ - Notice of Impoundment (MV2721)
- ☐ - Report to Superintendent (MV2722)

Unlicensed Prohibition:

- ☐ - Notice of Prohibition (MV2725)

Administrative Driving Prohibition:

- ☐ - Notice of Driving Prohibition (MV2685)
- ☐ - Report to Superintendent (MV2702a)
- ☐ - Certificate of Qualified Technician

Immediate Roadside Prohibition:

- ☐ - Notice of Driving Prohibition (MV2723)
- ☐ - Report to Superintendent (MV2724) & any attached narrative
- ☒ - Information relating to the calibration of the ASD

Other:

- ☐ - Missing fax pages: _____
- ☐ - Fax transmission illegible: _____

Any questions or concerns please call: James at 250-952-0380

OSMV USE ONLY

File#: s.22

* * * Communication Result Report (Aug. 22. 2012 9:12AM) * * *

1) osmv managment serv
2)

Date/Time: Aug. 22. 2012 9:08AM

File No. Mode	Destination	Pg(s)	Result	Page Not Sent
5215 Memory TX	fax server 812504597055	P. 1	OK OK	

Reason for error

E. 1) Hang up or line fail

E. 3) No answer

E. 5) Exceeded max. E-mail size

E. 2) Busy

E. 4) No facsimile connection

BRITISH
COLUMBIAMinistry of Justice
Office of the Superintendent of Motor Vehicles**URGENT**Enforcement Agency: ClintonFax: 250 459 7055 File number: s.22

The following documents (if they exist) are required in our office within 48 hours

**Please include this document with your
return fax to 250-978-8079**Vehicle Impoundment:

- ☐ - Notice of Impoundment (MV2721)
☐ - Report to Superintendent (MV2722)

Unlicensed Prohibitions:

- ☐ - Notice of Prohibition (MV2725)

Administrative Driving Prohibitions:

- ☐ - Notice of Driving Prohibition (MV2683)
☐ - Report to Superintendent (MV2702a)
☐ - Certificate of Qualified Technician

Immediate Roadside Prohibition:

- ☐ - Notice of Driving Prohibition (MV2723)
☐ - Report to Superintendent (MV2724) & any attached narrative
☒ - Information relating to the calibration of the ASD

Other:

- ☐ - Missing fax page(s)
☐ - Fax transmission illegible

Any questions or concerns please call: James at 250-962-0380

OSMV USE ONLY

File: s.22

This document is intended only for the use of the individual or organization to whom it is addressed, and it may contain confidential or personal information that is subject to provisions of the Access to Information Act and Privacy Act. This document must not be disclosed, copied or distributed to other unauthorized persons. If you have received this communication in error, please contact the sender immediately by telephone. Thank you.



Ministry of Justice
Office of the Superintendent of Motor Vehicles

URGENT

Enforcement Agency: Clinton

Fax: 250 459 7055 File number: s.22

The following documents (if they exist) are required in our office within 48 hours

**Please include this document with your
return fax to 250-978-8079**

Vehicle Impoundment:

- ☐ - Notice of Impoundment (MV2721)
- ☐ - Report to Superintendent (MV2722)

PLEASE SEE
ATTACHED REQ INFO.

Unlicensed Prohibition:

- ☐ - Notice of Prohibition (MV2725)

Administrative Driving Prohibition:

- ☐ - Notice of Driving Prohibition (MV2685)
- ☐ - Report to Superintendent (MV2702a)
- ☐ - Certificate of Qualified Technician

BML

CST. R. GOFFIC
CLINTON RCMP

Immediate Roadside Prohibition:

- ☐ - Notice of Driving Prohibition (MV2723)
- ☐ - Report to Superintendent (MV2724) & any attached narrative
- ☒ - Information relating to the calibration of the ASD

Other:

- ☐ - Missing fax pages: _____
- ☐ - Fax transmission illegible: _____

Any questions or concerns please call: James at 250-952-0380

OSMV USE ONLY

File#: s.22

The following material is intended only for the use of the individual or organization to whom it is addressed, and it may contain confidential or personal information that is subject to provisions of the *Freedom of Information and Protection of Privacy Act*. This material must not be distributed, copied or disclosed to other unauthorized persons. If you have received this transmission in error, please contact the sender immediately by telephone. Thank you.

ASD SERVICE EXPIRY DATE
2012/10/21

ASD SERIAL NUMBER
06778.3

CERTIFICATE OF QUALIFIED ASD CALIBRATOR

I, Andrew Butler, certify that:

I am a qualified ASD Calibrator, qualified to calibrate the
and (check one):

- ☒ Alco-Sensor IV DWF
☐ Alco-Sensor F5T

☒ On the 15th day of August, 2012 I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: Airgas Lot Number: A6101202, Expiry: 12 Jan 2013

It was found to be within the recommended limits.

OR

☐ On the _____ day of _____, 20____, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____; and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of September, 2012

Dated this 15th day of August, 2012 at Criston, in the Province of British Columbia

Calibrator: Andrew Butler

Dept/Reg Number: 54138

Agency: Ashcroft RCMP HP

[Signature]
Calibrator Signature

Page 26 redacted for the following reason:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2012/10/21

ASD SERIAL NUMBER
06778.3

CERTIFICATE OF QUALIFIED ASD CALIBRATOR

I, Andrew Butler, certify that:

I am a qualified ASD Calibrator, qualified to calibrate the

- ☒ Alco-Sensor IV DWP
☐ Alco-Sensor FST

and (check one):

☒ On the 15 day of August, 2012 I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: Airgas Lot Number: A649/202, Expiry: 12 Jan 2013

It was found to be within the recommended limits.

OR

☐ On the _____ day of _____, 20____, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____ and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of September 2012

Dated this 15 day of August, 2012 at Clinton, in the Province of British Columbia

Calibrator: Andrew Butler

Dept/Reg Number: 54138

Agency: Ashcroft RCMP HP

[Signature]
 Calibrator Signature

ASD ACCURACY CHECK I 3

Alc. Std.	(Dry Gas) Value @ Sea level	82 mg %
Info.	Manufacturer:	Airgas


Date	CALIBRATION CHECK						RECALIBRATION		CALIBRATION CHECK						BLANK TEST		ASD Calibrator Signature
	True-Cal Value	Calibration Check mg/100mL	True-Cal Value	Calibration Check mg/100mL	Calibration Check mg/100mL	Calibration Check mg/100mL	Calibration Check mg/100mL	True-Cal Value	# Samples taken	True-Cal Value	Calibration Check mg/100mL	Calibration Check mg/100mL	Calibration Check mg/100mL	Calibration Check mg/100mL	Blank Test mg/100mL	Calibration Expiry Date	
12/05/15	74	74	74	74	74	75			1						0	22.5.2012	Ally Hest
12/06/12	74	74	74	74	74	72			1							10.5.12	Ally Hest
12-07-19	74	74	74	74	74	74			1						0	16.4.2012	SA 54138
12/08/15	75	73	75	75	76				1						0	Sept 2012	SA 54138
12/09/13	75	61							1								OUT FOR SERVICE

Pages 29 through 32 redacted for the following reasons:

S.15, s.16, s.22

ASD Documentation Received - Fort Nelson RCMP

Thursday, February 28, 2013
11:44 AM

Subject	ASD Documentation Received - Fort Nelson RCMP
From	McKeown, Kayla JAG:EX
To	Esposito, Tony N JAG:EX; Anderson, Kathy E JAG:EX
Sent	Tuesday, October 23, 2012 4:46 PM
Attachments	 2012102315 4234.pdf

Hi Tony and Kathy,

Attached is a document received from Fort Nelson RCMP in relation to a police supplemental report regarding an ASD producing falsely high readings. I have checked through all of the processed IRP files from Fort Nelson and below is the information for the one IRP served that used the particular ASD in question. I will check through the files again later on this week. Please let me know if you require any other information.

Thanks!

ADP/IR P Prohibit ion Notice Number	ADP/IR P Offence Date	ADP/IR P Prohibit ion Type	ADP/IR P Prohibit ion Type 2	ADP/IR P Prohibit ion Type 3	ADP/IR P MVA Section	ADP/IR P Valid 7 or 30 Day IRP?	ADP/IR P No Impoun d on 3 or 7 Day IRP	ADP/IR P Police File Number	ADP/IR P Police Detach ment	ADP/IR P Police Region	ADP/IR P First ASD Result	ADP/IR P First ASD Serial Number	ADP/IR P Second ASD Result	ADP/IR P Second ASD Serial Number	VI Notice Number	VI Impoun d Date	VI Police File Number	VI Police Detach ment
s.22	20- Oct-201 2	IRP 90 day	IPR Fail/Ref use	IRP	ASD Fail		0	s.22	Fort Nelson (Northe rn Rockies)	4 - North RCMP	FAIL	39496	Not Responsive		s.22	20- Oct-201 2	s.22	Fort Nelson (Northe rn Rockies)

Kayla McKeown
Team Lead, Branch Support & Manual Operations
Administrative Justice Branch
The Office of the Superintendent of Motor Vehicles | Ministry of Justice
PO Box 9254 STN Prov Govt | Victoria BC | V8W 9J2 | 250-356-5634





REPORT TO SUPERINTENDENT SUPPLEMENTAL TO FILE

TO: OFFICE OF THE SUPERINTENDENT OF MOTOR VEHICLES
PO BOX 9254 STN PROV GOVT
VICTORIA, BC, V8W 9J2
FACSIMILE #: (250) 978-8079

DOCUMENT TYPE:

☒ IRP

IRP No: _____

☐ VI

VI No: _____

☐ UL

UL No: _____

☐ ADP

ADP No: _____

☐ 215 (24Hr)

215 No: _____

RE: DRIVER'S NAME (LAST, FIRST) _____

DOB: _____

BCDL: _____

POLICE FILE NO: _____

(YYYY/MM/DD)

DEPARTMENT/DETACHMENT:

FORT NELSON RCMP.

ISSUING OFFICER NAME:

Cst. A. ERICKSON

CONTACT PHONE:

250-774-2777

ATTENTION: Please add the following information to your records.

THE SUPPLEMENTAL REPORT IS:



DETAILED BELOW:

ASD #039496 showed to be producing falsely high readings on the calibration check done on Oct 23, 2012. The ASD was sent for the annual maintenance check which was due in Dec, 2012



ATTACHED BELOW: (Description of Document)

1. ASD accuracy log.

3. _____

2. _____

4. _____

CHECK LIST:



SUPPLEMENTAL REPORT COMPLETED AND/OR ATTACHED



ORIGINAL DOCUMENT ATTACHED

THIS REPORT CONSISTS OF 2 PAGES INCLUDING THIS PAGE

(Sworn or Solemnly Affirmed)

at FORT NELSON,

British Columbia,

on

YY

MM

DD

2012 10 23

before me,

A Commissioner for taking affidavits for British Columbia

Signature of Peace Officer

55740

PIN/Badge#

ASD ACCURACY CHECK LOG

Alc. Std. Info.	(Dry Gas) Value @ Sea Level	78 mg %	(Sim Sol'n) Expected value:	100 mg %
	Manufacturer:	ASTROGAS	Manufacturer:	CALWAVE
ASD SR: AS IV 08446		Simulator SR: 67561		
Location: Fort Meade BC		True-Cal SR: 008822		

Date	CALIBRATION CHECK				RECALIBRATION				CALIBRATION CHECK				BLANK TEST		Calibration Expiry Date	ASD Calibrator Signature
	Dry Gas Alc. Std. Conc. %	Cal Exp. Date	True-Cal Value	Calibration Check mg/100mL	Wei. Bait Alc. Std. Conc. %	Cal Exp. Date	# Sim Samples taken	True-Cal Value	Calibration Check mg/100mL	True-Cal Value	Calibration Check mg/100mL	Blank Test mg/100mL				
June 6/12	AG201502	14 Feb 2014			20723	2013/12	4	78	79	78	84					
June 6/12	AG201502	14 Feb 2014			20723	2013/12	5	78	78	78	79	0			July 3/12	
June 28/12	AG201502	14 Feb 2014	77 83		20723	2013/12	3	77	80	77	80	0			July 30/12	49990
July 13/12	AG201502	14 Feb 2014	78 82									0			Aug 23/12	AE 55740
Aug 17/12	AG201502	14 Feb 2014	78 86		20731	2014/04/30	1	78	82	78	82	0			Sept 14/12	AE 55740
Sept 14/12	AG201502	14 Feb 2014	77 80									0			Oct 11/12	12 49990
Sept 27/12	AG201502	14 Feb 2014	78 89		20731	2014/04/30	1	78	85						N/A	AE 55740
Sept 27/12	AG201502	14 Feb 2014			20731	2014/04/30	2	78	83	78	83	0			Oct 27/12	AE 55740
Oct 23/12	AG201502	14 Feb 2014	79 96													AE 55740

10/23/2012 04:33 FAX 12507742763

10/22/12 13:42:41 **SI** **VICTORIA OSMVOSMV**

Linked, stats. Corey

10/24/12 10:32:09 **JG** **VICTORIA OSMVOSMV**

due to highly false ratings on asd as stated in police supplemental report, this irp ^{s.22} is deemed invalid and cancelled as per team lead. contacted nelson police and they provided contact info for client. called client and stated that osmv has cancelled irp, will pay for towing and storage from oct 20.12 to oct 24.12 Vi^{s.22} faxed ilo order of release. sent police cpic notice.

ok to issue client free duplicate dl - jag

10/24/12 10:51:41 **JG** **VICTORIA OSMVOSMV**

correction - contacted fort nelson police - jag

10/24/12 11:58:45 **CG** **VICTORIA HEAD OFFICEHO**

Free dupe issued today. Cheryl ICBC

Pages 37 through 39 redacted for the following reasons:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2012-12-14

ASD SERIAL NUMBER
039496

CERTIFICATE OF QUALIFIED ASD CALIBRATOR

I, Allan Darcy ERICKSON, certify that:

I am a qualified ASD Calibrator, qualified to calibrate the
and (check one):

- ☒ Alco-Sensor IV DWF
☐ Alco-Sensor FST

☐ On the _____ day of _____, 20____, I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

OR

☒ On the 29 day of September, 2012, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: CALWAVE Lot Number: 20731, Expiry: 2014-04-30 and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: AIRGAS Lot Number: AG204502, Expiry: 14Feb2014.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of October 27, 2012.

Dated this 29 day of September, 2012, at Fort Nelson, in the Province of British Columbia

Calibrator: Allan Darcy ERICKSON

Dept/Reg Number: 55740

Agency: Fort Nelson RCMP



Calibrator Signature

Pages 41 through 46 redacted for the following reasons:

S.15, s.16, s.22

ASD Documentation Received - Sea to Sky Traffic Services/Squamish RCMP

Thursday, February 28, 2013
11:39 AM

Subject	ASD Documentation Received - Sea to Sky Traffic Services/Squamish RCMP
From	McKeown, Kayla JAG:EX
To	Esposito, Tony N JAG:EX; Anderson, Kathy E JAG:EX; Burchnell, Shelly K JAG:EX
Sent	Wednesday, July 18, 2012 4:00 PM
Attachments	 2012071815 5848

Hi Everyone,

Attached is a document received from Sea to Sky Traffic Services/Squamish RCMP in relation to a police supplemental report regarding an ASD. I have checked through all of the processed IRP files and below is the information for the 3 IRPs served that used the particular ASD in question. However, one of those IRPs served has already been cancelled for the reason that the Notice was served on an old IRP form. I will check through the files again later this week to ensure that all IRPs processed have covered all appropriate dates for the ASD.

Please let me know if you require any other information.

Thanks!

Date of Offence	Driver Name	DL#	Prov/St ate	Licence Class	Prohibiti on Type	IRP/ADP Notice #	First ASD Result	First ASD Serial #	Second ASD Result	Second ASD Serial #	VI Notice #	VI Offence	VI Length	Police File Number	Police Detach ment	Comme nts
2012-06-23	s.22		BC	s.22	3 day IRP - WARN	s.22	Not Responsive		WARN	055120		No Impound		s.22	Sea to Sky Traffic Services	Cancell ed as Notice was served on OLD IRP form
2012-06-24			ON		90 day IRP - FAIL		FAIL	055120	NOT REQUE STED		s.22	30 day impoun dment (30 or 90 day IRP)	30 day		Sea to Sky Traffic Service s	s.22
2012-07-05			BC		3 day IRP - WARN		FAIL	055120	Not Responsive			3 day impoun dment (3 day IRP)	3 day		Sea to Sky Traffic Service s	

Kayla McKeown
Team Lead, Branch Support & Manual Operations
Administrative Justice Branch
The Office of the Superintendent of Motor Vehicles | Ministry of Justice
PO Box 9254 STN Prov Govt | Victoria BC | V8W 9J2 | 250-356-5634



Pages 48 through 52 redacted for the following reasons:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2012-12-19

ASD SERIAL NUMBER
055120

CERTIFICATE OF A QUALIFIED ASD CALIBRATOR

I, Cst M.R. JANSON, certify that

I am a qualified ASD Calibrator, qualified to calibrate the

☒ Alco-Sensor IV DWF

☐ Alco-Sensor FST

- ☐ On the _____ day of _____, 20____, I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

- ☒ On the 19 day of June, 2012, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: Calwave Lot Number: 80141, Expiry: Aug. 2013 and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: Airgas Lot Number: AG114604, Expiry: 26 May 2013.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.


To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of July 17th, 2012.

Dated this 19 day of June, 2012, at Squamish, in the Province of British-Columbia

Calibrator: Martin Janson

Dept/Reg Number: 50778

Agency: RCMP


Calibrator Signature



July 23, 2012

s.22

Immediate Roadside Prohibition (IRP) No. s.22

I am writing with respect to the Notice of Driving Prohibition served on you on June 24, 2012.

Please be advised that the Notice is deemed incomplete, and therefore, the driving prohibition and its related vehicle impoundment have been cancelled. Your requirement for the Responsible Driver Program and the Ignition Interlock Program has also been cancelled.

The registered owner of the vehicle will be notified by this office that the Superintendent will pay the towing and storage up to and including this date.

Appeals Registry
Office of the Superintendent of Motor Vehicles

/km

Pages 55 through 58 redacted for the following reasons:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2012-12-19

ASD SERIAL NUMBER
055120

CERTIFICATE OF A QUALIFIED ASD CALIBRATOR

I, Cst M.R. JANSON, certify that

I am a qualified ASD Calibrator, qualified to calibrate the

☒ Alco-Sensor IV DWF

☐ Alco-Sensor FST

- ☐ On the _____ day of _____, 20____, I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

- ☒ On the 19 day of June, 2012, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: Calwave Lot Number: 80141, Expiry: Aug. 2013 and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: Airgas Lot Number: AG114604, Expiry: 26 May 2013.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of July 17th, 2012.

Dated this 19 day of June, 2012, at Squamish, in the Province of British-Columbia

Calibrator: Martin Janson

Dept/Reg Number: 50778

Agency: RCMP



Calibrator Signature

Pages 60 through 61 redacted for the following reasons:

S.15, s.16, s.22



July 23, 2012

s.22

Immediate Roadside Prohibition (IRP) No. s.22

I am writing with respect to the Notice of Driving Prohibition served on you on July 5, 2012.

As the Notice is incomplete, the driving prohibition and its related vehicle impoundment have been cancelled and removed from your driving record.

Your vehicle was due for release on July 9, 2012. Upon receipt of your proof of payment, the Superintendent will pay towing and storage costs up to and including this date. Please remit a copy of this letter, the Impound Lot invoice and your copy of receipt of payment to this office via the address or fax number below. You are responsible for storage costs beyond this date.

I note that you reinstated your driver's licence associated with this driving prohibition. A refund of this fee has been authorized and will be sent to you in approximately six to eight weeks. If you have any outstanding fines or debts to the Province of British Columbia, your refund may be used to help offset those debts.

Intake Services
Office of the Superintendent of Motor Vehicles

/km

Pages 63 through 68 redacted for the following reasons:

S.15, s.16, s.22

ASD SERVICE EXPIRY DATE
2012-12-19

ASD SERIAL NUMBER
055120

CERTIFICATE OF A QUALIFIED ASD CALIBRATOR

I, Cst M.R. JANSON, certify that

I am a qualified ASD Calibrator, qualified to calibrate the

☒ Alco-Sensor IV DWF

☐ Alco-Sensor FST

- ☐ On the ____ day of _____, 20____, I checked the calibration of the ASD with the above noted serial number by means of a dry gas Alcohol Standard, Manufacturer: _____ Lot Number: _____, Expiry: _____.

It was found to be within the recommended limits.

- ☒ On the 19 day of June, 2012, I recalibrated the ASD with the above noted serial number by means of a wet bath Alcohol Standard, Manufacturer: Calwave Lot Number: 80141, Expiry: Aug. 2013 and checked the calibration by means of a dry gas Alcohol Standard, Manufacturer: Airgas Lot Number: AG114604, Expiry: 26 May 2013.

It was found to be within the recommended limits.

AND

The ASD calibration was conducted in accordance with the training that I received.

The Alcohol Standard(s) used to check and/or calibrate the ASD was/were suitable for this purpose.

To the best of my knowledge the ASD is functioning correctly. The ASD has a calibration expiry date of July 17th, 2012.

Dated this 19 day of June, 2012, at Squamish, in the Province of British-Columbia

Calibrator: Martin Janson

Dept/Reg Number: 50778

Agency: RCMP



Calibrator Signature