


[illegible]

INVOICE

smiths detection

bringing technology to life

Smiths Detection Montreal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ph No - 905-817-5990
<http://www.smithsdetection.com>



Bill to No Bill to name S.17 FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 -256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	INVOICE 93028810 Date 06/13/2011
Phone No 604-462-9313	Sales Order No. 93753 Date 16-MAY-2011
Customer Contact PAT GAMACHE	Customer Order No. Warranty repair Date 06/07/2011
Bill to VAT reg.No. Customer Contact PAT GAMACHE	Sold to No. S.17 FRASER REGIONAL CORR CENTRE
Ship To Ship to name S.17 FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 -256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Payment Terms Z009 Net 30
Customer Contact PAT GAMACHE Phone No 604-462-9313 Fax No EIN No.	Due date 07/13/2011
	Our Contact Person
	Email
	Our Contact2


Item No	Material Information	Qty	UoM	Unit Price	Total Price	Curr
010400	Material No 40090 Material Description IONSCAN 400B DEPOT REPAIR CA35 Country of Origin Serial no 19514	1.000	EA	0.00	0.00	CAD

Gross Value	0.00	CAD
Total Net Value	0.00	CAD
Grand Total	0.00	CAD

Delivery Note

smiths detection

bringing technology to life




Ship to S.17 FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada		Delivery No. 80135610
Ship to Name FRASER REGIONAL CORR CENTRE		Date 06/07/2011
Ship to Address MIN OF PUBLIC SAFETY & SOL GEN 13777 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada		Sales Order No. 93753
EIN No. PAT GAMACHE		Date 06/18/2011
Customer Contact PAT GAMACHE		Customer Order No Warranty repair
Phone No. 604-462-9313		Date 06/07/2011
Fax No. 604-462-9313		

Item No.	Material Information	Qty	UoM	Net Weight	UoM
10	Order Material No. Order Material Description Serial No. Sales Order Qty Yel to Delivery Qty	4810788POS DETECTOR MODULE, NARC MODE, 400B 19514 1 0	1	EA	0.000
					KG

Terms of Delivery Mode of Shipment Name of Forwarder Transport Reference No.	EXW TORONTO
---	-------------

smiths detection
bringing technology to life

Repair Summary Form



Date (mm/dd/yy) May 31/2011

Serial Number	15734
Model Type	4008

Problems Reported :

Waiting for flow

Flow rate on status reading 0 cc/min

Problems found during inspection (Verify Customer Complaint; note any other issues):

Nafion tube has leak.

Problems / Notes during repair:

Nafion Tube repaired.

Unit tested.

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montreal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ph No - 905-817-5990
<http://www.smithsdetection.com>

00010001

Bill to No	S.17	INVOICE	93028389	Date	06/02/2011
Bill to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Sales Order No.	93755	Date	18-MAY-2011
Bill to VAT reg. No.		Customer Order No.	Warranty repair	Date	05/31/2011
Customer Contact	PAT GAMACHE	Sold to No.	S.17		
Phone No	604-462-9313	Sold to name	FRASER REGIONAL CORR CENTRE		
Fax No		Payment Terms	2009	Net 30	
Ship To	S.17	Due date	07/02/2011		
Ship to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Our Contact Person			
Customer Contact	PAT GAMACHE	Email			
Phone No	604-462-9313	Our Contact2			
Fax No					
EIN No.					

Item No	Material Information	Qty	UoM	Unit Price	Total Price	Curr
010400	Material No Material Description Country of Origin Serial no	40080 IONSCAN 4008 DEPOT REPAIR CA35 15734	1.000 EA	0.00	0.00	CAD

Gross Value	0.00	CAD
Total Net Value	0.00	CAD
Grand Total	0.00	CAD

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montreal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ptn No - 805-817-5990
<http://www.smithsdetection.com>

Bill to No S.17	INVOICE	93025405	Date	03/29/2011
Bill to name FRASER REGIONAL CORR CENTRE MAIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Sales Order No.	8333B	Date	22-FEB-2011
Bill to VAT reg No	Customer Order No.	Pay by Mastercard	Date	03/28/2011
Customer Contact PAT GAMACHE	Sold to No. S.17	FRASER REGIONAL CORR CENTRE		
Phone No 604-462-9313				
Fax No	Payment Terms	2009	Net 30	
Ship To S.17	Due date	04/28/2011		
FRASER REGIONAL CORR CENTRE MAIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Our Contact Person			
Customer Contact PAT GAMACHE	Email Our Contact			
Phone No 604-462-9313				
Fax No				
Bill to No				

Item No	Material Information	Qty	UoM	Unit Price	Total Price	Curr
010300	Material No Material Description Country of Origin Serial no 19514	1.000	EA	2,300.00	2,300.00	CAD
010400	Material No Material Description Country of Origin Serial no 1810889-F PUMP BRUSHLESS, 24V MOD PUMP US	1.000	EA	513.70	513.70	CAD
010500	Material No Material Description Country of Origin Serial no 7049704 FILTER, SAMPLE DISPOSABLE CLRN NYLON US	1.000	EA	19.00	19.00	CAD
010600	Material No Material Description Country of Origin Serial no 1810918-A SURGE TANK TECFA CA	1.000	EA	123.00	123.00	CAD
010700	Material No Material Description Country of Origin Serial no 40128 IONSCAN 400B LABOUR CA35	1.000	EA	405.00	405.00	CAD
010800	Material No Material Description Country of Origin Serial no 40320 IONSCAN 400B DEPOT REPAIR CA35	1.000	EA	0.00	0.00	CAD

Gross Value	3,360.70	CAD
Total Net Value	3,360.70	CAD
Total Before Tax	3,360.70	CAD

[Signature]
SCANNED

7650

03/31/2011
BKD 11336 6938 1500000

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montreal Inc
7930 Century Avenue
Mississauga, Ontario, L6N 2V8
Canada
Ph No - 905-817-5990
<http://www.smithsdetection.com>

Bill to No	S 17	INVOICE	93025405	Date	03/29/2011
Bill to Name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 MAPLE RIDGE British Columbia V2X 7G3 Canada				

Taxable AMT	3,360.70	CAD
Total Tax	403.29	CAD
Grand Total	3,763.99	CAD

Repair Summary Form

Date (mm/dd/yy)	03/28/11
-----------------	----------

Serial Number	19514
Model Type	400B

Problems Reported :

No flow

Problems found during inspection (verify Customer Complaint; note any other issues):

Vacuum pump failing

Problems / Notes during repair:

Failing vacuum pump replaced,

Blocked surge tank and also filter replaced.

Im's was blocked repaired.

smiths detection

bringing technology to life

CERTIFICATE NO.: B030311- 8

7030 Century Ave.
Mississauga, Ontario
Canada, L5N 2V8
Phone: (905) 817-5990
Fax: (905) 817-5992

RADIATION LEAK TEST CERTIFICATE

Part 1- Leak Test Sample Collection

Company name: Smiths Detection License #: 14431-1-15
Address: 7030 Century Ave., Mississauga, ON L5N 2V8
Phone No.: (905) 817-5990 Fax No.: (905) 817-5992
Date taken: March 3/2011 Contact person: Paul Lynds cell: (416) 729-8048
Taken by: Neil Gillis Signature: *Neil Gillis*
Type of source: Nickel-63 Nominal activity: 555 MBq (15.00 mCi)
Type of sample: Q-tip cotton swab
System serial No.: 15734
IMS serial No. (1): 400B-0088 Source serial No. (1): G-667
(eg. SAB, PUMS, 400B, 500DT) (eg. G, KI, KU)
IMS serial No. (2): _____ Source serial No. (2): _____
(eg. SAB, PUMS, 400B, 500DT) (eg. G, KI, KU)
Collection method: Wipe test is taken from exhaust port, then from IMS detector inlet.

Part 2- Leak Test Sample Measurement

Background counts / minute: 25 Sample counts / minute: 95
Net measured radioactivity: 2.4 Bq. Passed ☒ Failed* ☐
Test Performed By: Jasvir Mann Date: Mar 03/11
Signature: *Jasvir Mann*
Test device: Protein Instruments, alpha/beta counter ☐ IPC-9025 ☒ MPC-9400
(calibrated daily)
* Rejectable Upper Limit: 200 Bq. (5.4 nCi)
In case of failure, notify contact listed above, quarantine device, submit second sample.
1 Becquerel (Bq) = 1 disintegration per second (dps) = 27 picocuries (pCi)

Emergency Contact: Paul Lynds, cell phone: (416) 729-8048

CNSC Licence #: 14431 - 1-15.0

For Instruments Imported to Japan

Test method: Wipe test defined on Manuf. Instructions - PS100145proc

Japanese Design Certification Number: ☐ 010 ☒ 011 ☐ 019 ☒ 127
(400B Series) (SABre Series) (500DT) (IMATD)

FORM156

September, 27, 2010, Rev.: 1

7030 Century Ave.
Mississauga, Ontario
Canada, L5N 2V8
Phone: (905) 817-5990
Fax: (905) 817-5992

CERTIFICATE NO.: 8023811 3

RADIATION LEAK TEST CERTIFICATE

Part 1 - Leak Test Sample Collection

Company name: Smiths Detection License #: 14431-1-15

Address: 7030 Century Ave., Mississauga, ON, L5N 2V8

Phone No.: (905) 817-5990 Fax No.: (905) 817-5992

Date taken: March 28, 2011 Contact person: Paul Lynds, cell: (416) 729-8048

Taken by: Dilek Sezgin Signature: [Signature]

Type of source: Nickel-63 Nominal activity: 1555 MBq (15.00 mCi)

Type of sample: Q-tip cotton swab

System serial No.: 19514

IMS serial No. (1): 400B-8797 Source serial No. (1): 00964
(eg. SAB, PMS, 400B, 500DT) (eg. SAB, PMS, 400B, 500DT) (eg. G, KT, KU)

IMS serial No. (2): _____ Source serial No. (2): _____
(eg. SAB, PMS, 400B, 500DT) (eg. G, KT, KU)

Collection method: Wipe test is taken from exhaust port, then from IMS detector inlet.

Part 2 - Leak Test Sample Measurement

Background counts / minute: _____ Sample counts / minute: 64

Net measured radioactivity: _____ Bq. Passed [☒] Failed* [☐]

Test Performed By: _____ Date: March 28, 2011

Signature: _____

Test device: Protec Instruments: alpha/beta counter [☐] PC-9025 [☐] MPC-9400
(Calibrated daily)


* **Rejectable Upper Limit: 200 Bq. (5.4 nCi)**
In case of failure, notify contact listed above, quarantine device, submit second sample.
1 Becquerel (Bq) = 1 disintegration per second (dps) = 27 picocuries (pCi)

Emergency Contact: Paul Lynds, cell phone: (416) 729-8048 CNSC Licence #: 14431 - 1-15.0

For Instruments Imported to Japan

Test method: Wipe test defined on Manuf. Instructions - PS100145proc

Japanese Design Certification Number: [☐] 010 [☐] 011 [☐] 019 [☐] 127
(400B Series) (Sabre Series) (500DT) (MMTD)



Repair Summary Form

Date (mm/dd/yy) 03/14/11

Serial Number	15734
Model Type	400B

Problems Reported :

No Flow

Problems found during inspection (Verify Customer Complaint; note any other issues):

IMS leaking 256cc, Turned unit on and drift flow 0cc, Desorber has black substance on it and smells when heated.

Problems / Notes during repair:

IMS Rework, Found leak in drierite disconnect fitting, Replace Desorber. Clean slide

Calibrate IMS and flow module, Run solutions, Upgrade CP's and FW

smiths

bringing technology to life

CERTIFICATE NO.: B030311-8

7030 Century Ave.
Mississauga, Ontario
Canada, L5N 2V8
Phone: (905) 817-5890
Fax: (905) 817-5992

RADIATION LEAK TEST CERTIFICATE

Part 1 - Leak Test Sample Collection

Company name: Smiths Detection License #: 14431-1-15
Address: 7030 Century Ave., Mississauga, ON, L5N 2V8
Phone No.: (905) 817-5990 Fax No.: (905) 817-5992
Date taken: March 3/2011 Contact person: Paul Lynds cell: (416) 729-8048
Taken by: Neil Gillis Signature: *Neil Gillis*
Type of source: Nickel-63 Nominal activity: 555 MBq (15.00 mCi)
Type of sample: Q-tip cotton swab
System serial No.: 15734
IMS serial No. (1): 400B-0088 Source serial No. (1): G-667
(eq. SAB + PIMS, 400B, 500DT) (eq. G, KI, KU)
IMS serial No. (2): Source serial No. (2):
(eq. SAB + PIMS, 400B, 500DT) (eq. G, KI, KU)
Collection method: Wipe test is taken from exhaust port, then from IMS detector inlet.

Part 2 - Leak Test Sample Measurement

Background counts / minute: 95 Sample counts / minute:
Net measured radioactivity: Bq. Passed ☒ Failed* ☐
Test Performed By: Jasvir Mann Date: March 3/11
Signature: *Jasvir Mann*
Test device: Prolean Instruments: alpha/beta counter ☐ IPC-9025 ☒ MPC-9400
(Calibrated daily)
* **Rejectable Upper Limit: 200 Bq. (5.4 nCi)**
In case of failure, notify contact listed above, quarantine device, submit second sample.
1 Becquerel (Bq) = 1 disintegration per second (dps) = 27 picocuries (pCi)
Emergency Contact: Paul Lynds, cell phone: (416) 729-8048 CNSC Licence #: 14431 - 1 - 15.0

For Instruments Imported to Japan

Test method: Wipe test defined on Manuf. Instructions - PS100145proc
Japanese Design Certification Number: ☒ 010 ☒ 011 ☒ 019 ☒ 127
(400B Series) (Sabre Series) (500DT) (MMTD)

FORM166

September, 27, 2010, Rev.: 1

Delivery Note



smiths detection
 bringing technology to life
 Smiths Detection Montreal Inc
 7030 Century Avenue
 Mississauga Ontario L5N 2V8
 Canada
 Pn No. - 905 817 5990
 Fax No. - 905 817 5992
 Website - <http://www.smithsdetection.com>

Ship to S.17 Ship to Name FRASER REGIONAL CORR CENTRE Ship to Address MIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada		Delivery No. 80122068 Date 03/23/2011 Sales Order No. 83332
EIN No. Customer Contact PAT GAMMACHE Phone No. 604-462-9313 Fax No.		Date 02/22/2011 Customer Order No Pay by Mastercard Date 03/23/2011

Item No.	Material Information	Qty	Uom	Net Weight	Uom
10	Order Material No. 4810768-AZ Order Material Description DETECTOR MODULE, 400B, BASIC Serial No. 15734 Sales Order Qty 1 Yct to Delivery Qty 0	1	EA	0.000	KG

Terms of Delivery Mode of Shipment Name of Forwarder Transport Reference No.	EXW TORONTO
---	-------------

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montreal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ph No - 905-617-5990
<http://www.smithsdetection.com>

Bill to No Bill to name	S.17 FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	INVOICE Sales Order No.	93025287	Date	03/26/2011
Bill to VAT reg. No.		Customer Order No.	83332	Date	22-FEB-2011
Customer Contact	PAT GAMACHE	Pay by Mastercard		Date	03/23/2011
Phone No	604-462-9313	Sold to No.	S.17		
Fax No		Sold to name	FRASER REGIONAL CORR CENTRE		
Ship To	S.17 FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Payment Terms	Z009	Net 30	
Ship to name		Due date	04/24/2011		
Customer Contact	PAT GAMACHE	Our Contact Person			
Phone No	604-462-9313	Email			
Fax No		Our Contact2			
EIN No.					

Item No	Material Information	Qty	Unit	Price	Total Price	Currency
010300	Material No Material Description Country of Origin Serial no	1.000	EA	2,300.00	2,300.00	CAD
	4810769RW IMS ASSY,(REWORK)		PST		161.00	
010400	Material No Material Description Country of Origin Serial no	1.000	EA	1,576.84	1,576.84	CAD
	2814989-D DESORBER ASSY 400B		PST		110.38	
010500	Material No Material Description Country of Origin Serial no	1.000	EA	270.00	270.00	CAD
	40138 ION IONSCAN 400B LABOUR CA35		PST		18.90	
010700	Material No Material Description Country of Origin Serial no	1.000	EA	0.00	0.00	CAD
	40080 IONSCAN 400B DEPOT REPAIR CA35					
	15734 Serial no					
	15734 Material Description					
	15734 Country of Origin					

Gross Value	4,146.84	CAD
Total Net Value	4,146.84	CAD
Total Before Tax	4,146.84	CAD
Taxable AMT	4,146.84	CAD
Total Tax	497.62	CAD
Grand Total	4,644.46	CAD

APPROVED FOR PAYMENT

SIGNATURE

March 30/11

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montréal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ph No - 905-817-5990
<http://www.smithsdetection.com>



Bill to No	S.17	INVOICE	93019667	Date	11/17/2010
Bill to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 -256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Sales Order No.	70374	Date	28-OCT-2010
Bill to VAT reg.No.		Customer Order No.	Pay by Mastercard		
Customer Contact	PAT GAMACHE	Sold to No.	S.17		
Phone No	604-462-9313	Sold to name	FRASER REGIONAL CORR CENTRE		
Fax No					
Ship To	S.17	Payment Terms	Z009	Net 30	
Ship to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL.GEN 13777 -256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Due date	12/17/2010		
Customer Contact	PAT GAMACHE	Our Contact Person			
Phone No	604-462-9313	Email			
Fax No		Our Contact2			
EIN No.					

Item No	Material Description	Material No	15030	CONDENSER TUBE ASSEMBLY* DISPOSABLE	1.000	EA	PST	53.00	53.00	CAD
010300	Country of Origin	CA							3.71	
010400	Serial no	15734								
	Material No	7043704							38.00	CAD
	Material Description	FILTER, SAMPLE DISPOSABLE CLR NYLON							2.66	
	Country of Origin	US								
010500	Serial no	15734								
	Material No	3813844-B							183.03	CAD
	Material Description	GRAPHICS PCB ASSY 400B							12.81	
	Country of Origin	CA								
010600	Serial no	15734								
	Material No	40138							540.00	CAD
	Material Description	ION IONSCAN 400B LABOUR CA25							37.80	
	Country of Origin	ION								
010800	Serial no	15734								
	Material No	40090							0.00	CAD
	Material Description	IONSCAN 400B DEPOT REPAIR CA35								
	Country of Origin	IONSCAN								
	Serial no	15734								

Gross Value	814.03	CAD
Total Net Value	814.03	CAD
Total Before Tax	814.03	CAD
Taxable AMT	814.03	CAD
Total Tax	97.68	CAD
Grand Total	911.71	CAD

INVOICE

smiths detection
bringing technology to life

Smiths Detection Montreal Inc
7030 Century Avenue
Mississauga, Ontario, L5N 2V8
Canada
Ph No - 905-817-5990
<http://www.smithsdetection.com>

Bill to No	S.17	INVOICE	93018575	Date	10/21/2010
Bill to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Sales Order No.	65736	Date	20-SEP-2010
Bill to VAT reg.No.		Customer Order No.	Pay by Mastercard	Date	10/18/2010
Customer Contact	PAT GAMACHE	Sold to No.	S.17		
Phone No	604-462-9313	Sold to name	FRASER REGIONAL CORR CENTRE		
Fax No					
Ship To	S.17	Payment Terms	Z009	Net 30	
Ship to name	FRASER REGIONAL CORR CENTRE MIN OF PUBLIC SAFETY & SOL GEN 13777 - 256TH ST. BOX 1500 MAPLE RIDGE British Columbia V2X 7G3 Canada	Due date	11/20/2010		
Customer Contact	PAT GAMACHE	Our Contact Person			
Phone No	604-462-9313	Email			
Fax No		Our Contact2			
EIN No.					

Item No	Material Information	Qty	UoM	Unit Price	Total Price	Curr
010300	Material No Material Description Country of Origin Serial no 19514 40138 ION IONSCAN 400B LABOUR CA35	1.000	EA	2,300.00 PST	2,300.00 161.00	CAD
010400	Country of Origin Material No Material Description Country of Origin Serial no 18514 40090 IONSCAN 400B DEPOT REPAIR CA35	1.000	EA	405.00 PST	405.00 28.35	CAD
010700	Material No Material Description Country of Origin Serial no 19514	1.000	EA	0.00	0.00	CAD

Gross Value	2,705.00	CAD
Total Net Value	2,705.00	CAD
Total Before Tax	2,705.00	CAD
Taxable AMT	2,705.00	CAD
Total Tax	324.60	CAD
Grand Total	3,029.60	CAD